



# Clip Slings

---




**Unpadding clip sling**

The above has no padding in the leg area and therefore is easier to apply to residents with skeletal issues such as contractures and curvatures of the spine. These slings are commonly used on people who are in moulded wheelchairs.




**Padded sling**

Padding adds a level of comfort to the leg area of the slings. The Padding also helps to prevent the material roping and creasing under legs.



**Mesh clip sling**

The above is most commonly used for showering or bathing. The mesh allows the water to pass through the material therefore retaining less water when the patient is being towel dried. This is a benefit when the patient is hoisted back into their wheelchair or onto a towel in bed. The porous material also allows the skin to breath and can be left underneath the patient.




**Toilet clip sling**

Used for toileting, dressing and undressing patients. The padded arm supports provide the comfort and support. Leg supports are padded for comfort but again are narrower to allow clothing to be removed. Careful patient assessment is necessary before this sling is used.


# Clip Slings

---




## Extended leg padded clip sling

The leg pieces are extended by 100mm (4 inches) to allow easier attachment of the leg clips. Used for patients who have larger thighs and hips.




## Amputee clip sling with Velcro flap

Used on double amputee above or below knee. The design allows easier application to the patient rather than having to log roll them in their chair or bed. The sling is applied as with a normal sling and the Velcro fastened to provide more support.



## Right leg single amputee clip sling

Used for Left single leg amputee patients above, below or total hip. The larger leg piece provides more support to the stumps of the affected side.




## Left leg single amputee clip sling

Used for Right single leg amputee patients above, below or total hip. The larger leg piece provides more support to the stumps of the affected side.


# Clip Slings

---



## **Amputee clip sling**

Used for Bilateral leg amputee patients with above knee or total hip. The design requires the patient to be log rolled onto the sling while on the bed or rolled from side to side in a chair. The aperture does not allow toileting.




## **Amputee clip sling with commode aperture**

Used for Bilateral leg amputee patients with above knee or total hip. The design requires the patient to be log rolled onto the sling while on the bed or rolled from side to side in a chair. The aperture allows toileting of a patient who's lower clothing has already been removed while on the bed.


# Standing and Raising slings- with clip fixing

Position the sling around the patient's back so that it lies 50mm (2") or so, horizontally above the patients' waistline, with the patients' arms outside the sling. Ensure the support strap is separated, brought loosely around the body, and is not twisted or trapped behind the patients' back.

Fasten the support strap securely; the strap should be tight, but comfortable for the patient. (See fig. 1). The support strap will assist in supporting the patient in the sling during the lifting procedure.




Bring the Lifter carefully up to the patient, placing the patient's feet on the footrest, and continue forward until the kneepad is just in contact with the patient's knees and upper shin. Put on the brakes. (See fig.2).



Lower the support arms carefully, using the handset control, until the sling attachment lugs on the support arms are close enough to connect up to one

of the sling attachment clips. (See fig. 3).




Next, select an appropriate clip on the opposite end of the sling, and connect up.

- **Note:** The patient should be supported but not pulled forward too much.

Ensure the sling clips are pulled well down onto the attachment lugs and secure, before attempting to lift the patient.

If possible, the patient should now hold onto the grab handles with one or both hands. (See fig. 4).






# Clip Slings


## STANDARD SLING

Order ref	Size
MAA2000-XS	Extra Small
MAA2000-S	Small
MAA2000-M	Medium
MAA2000-L	Large
MAA2000-LL	Large Large
MAA2000-XL	Extra Large
MAV2908	Extra Extra Large




## MESH SLING

Order ref	Size
MAA4060-XS	Extra Small
MAA4060-S	Small
MAA4060-M	Medium
MAA4060-L	Large
MAA4060-LL	Large Large
MAA4060-XL	Extra Large
MAV2994	Extra Extra Large




## EXTENDED LEG SLING + 100MM

Order ref	Size
MAV4841	Small
MAA4100-M	Medium
MAA4100-L	Large
MAA4100-XL	Extra Large
MAA4100-XXL	Extra Extra Large




## EXTENDED LEG UNPADDED SLING + 100MM

Order ref	Size
MAA2100-M	Medium
MAA2100-L	Large
MAA2100-XL	Extra Large




## RIGHT LEG SINGLE AMPUTEE SLING

Order ref	Size
MAV4826	Medium
MAV4662	Large
MAV5767	Large Large
MAV4820	Extra Large



## DOUBLE AMPUTEE SLING

Order ref	Size
MAA2090-XS	Extra Small
MAA2090-S	Small
MAA2090-M	Medium
MAA2090-L	Large
MAA2090-XL	Extra Large




## DOUBLE AMPUTEE MESH SLING

Order ref	Size
MAV2596	Small
MAV2597	Medium
MAV2598	Large
MAV2599	Extra Large




## PADDED SLING

Order ref	Size
MAA4000-XS	Extra Small
MAA4000-S	Small
MAA4000-M	Medium
MAA4000-L	Large
MAA4000-LL	Large Large
MAA4000-XL	Extra Large
MAV4834	Extra Extra Large




## TOILET SLING WITH HEAD SUPPORT

Order ref	Size
MAA4031-XS	Extra Small
MAA4031-S	Small
MAA4031-M	Medium
MAA4031-L	Large
MAA4031-LL	Large Large
MAA4031-XL	Extra Large



## AMPUTEE SLING WITH VELCRO FLAP

Order ref	Size
MAV4828	Extra Small
MAV4830	Small
MAV4818	Medium
MAV4673	Large
MAV5766	Large Large
MAV5050	Extra Large




## EXTENDED LEG MESH SLING + 100MM

Order ref	Size
MAA4160-M	Small
MAA4160-L	Medium
MAA4160-XL	Extra Large




## LEFT LEG SINGLE AMPUTEE SLING

Order ref	Size
MAV4873	Medium
MAV4509	Large
MAV5768	Large Large
MAV4867	Extra Large



## DOUBLE AMPUTEE SLING WITH COMMODO APERTURE

Order ref	Size
MAV7121	Small
MAV2630	Medium
MAV2632	Large
MAV2633	Extra Large
MAV7122	Extra Extra Large



## DOUBLE AMPUTEE MESH WITH COMMODO SLING

Order ref	Size
MAV7136	Small
MAV7040	Medium
MAV7076	Large
MAV7137	Extra Large

