



Pump	ePump	Joey	Connect
SKU	382400	383400	384400
Applications	Enteral Feed and Flush Pump with Pole Clamp	Enteral Feed and Flush Pump with Pole Clamp	Portable Enteral Feeding Pump
Type Infusion Device	Volumetric	Volumetric	Volumetric Enteral
Pumping Mechanism	Rotary Peristaltic	Rotary Peristaltic	Rotary Peristaltic
Service Life (years)	NA	5	5
Pump Shelf Life (years)	NA	5	5

Size	ePump	Joey	Connect
Dimensions (w x h x d, inch)	6.4 x 6.6 x 4.6	5.1 x 4.1 x 3.6	6.1 x 3.9 x 1.6
Total Dimensions (cubic inch)	194.304	75.276	38.064
Weight (lbs.)	2.3	1.7	0.73

Performance	ePump	Joey	Connect
Feeding Formula Delivery Rate	1-400 mL/hr	1-400 mL/hr	1-600 ml/hour
Priming Rate	NA	NA	1200 mL/hour
Formula Dose	NA	NA	1-3000 mL
Bolus Volume	1-3000 mL	1-3000 mL	1-3000 mL
Number of Boluses	Jan-99	Jan-99	Jan-99
Bolus Interval (hours)	1-24	1-24	1-24
Accuracy	±10% or 0.5 mL/hr	7% or 0.5 mL/hr	5% or 0.5 mL/hr
Occlusion Pressure	15 psi nominal	15 psi nominal	20 psi maximum
Flush Interval (hours)	1-24	1-24	NA
Flush Volume	10-500	10-500	NA
Auto Prime (Yes/No)	Yes	Yes	Yes
Continuous Feed (Yes/No)	Yes	Yes	Yes
Intermittent Feed (Yes/No)	Yes	Yes	No
Settings Lock (Yes/No)	No	No	Yes
Cassette Loading (Yes/No)	NA	NA	Yes

Operation	ePump	Joey	Connect
Power Input Requirements	NA	NA	100-240 V~, 50-60 Hz, 1.0A (1.0 A – 0.5 A)
Power Output	9V DC, 1.5 Amps	5V DC, 2.4 A	+5 V, 4.0 A
Battery	NIMH Battery Pack, Rechargeable	Lithium Ion Battery Pack, Rechargeable	Lithium Ion, Rechargeable
Battery Duration	≥ 15 hrs at 100 mL/hr feed rate	18 hrs at 125 mL/hr feed rate	24 hr @ 125 mL feed
Battery Recharge Time (hours)	8	6	7
Alarm Volume (maximum dBA)	70	65	65
Memory (feeding history)	72 hours	72 hours	72 hours
Pump Set Compatibility	Kangaroo™ ePump MISTIC Feed-Only Set or Feed & Flush Set	Kangaroo™ Joey enteral feeding pump sets	Kangaroo™ Connect feeding sets
Patient Use	Adults	Adults	Adults and Pediatric
Maximum Altitude for Operation (feet above sea level)	NA	6,561	13,123
Alarms	System Error Flow Error Pump Set Dislodged Error Hold Error Rotor Error Feed Error Flush Error Battery Low	System Error Flow Error Pump Set Dislodged Error Hold Error Rotor Error Feed Error Flush Error Battery Low	System Error Feed Bag Empty Supply Tube Blocked Patient Tube Blocked Cassette Dislodged Cassette Error Rotor Stuck Dead Battery Feeding Complete Feeding Incomplete Low Battery Pump Inactive
Notifications	NA	NA	

Additional Information	ePump	Joey	Connect
Backpack (Yes/No)	No	Optional - Mini or Super Mini	Yes
Integrated Handle (Yes/No)	Yes	No	No
Tabletop Use (Yes/No)	Yes	Yes	Yes
IV Pole Mounting (Yes/No)	Yes	Yes	Yes
Bed Rail Mounting (Yes/No)	No	No	Yes
DEHP free (Yes/No)	NA	Yes	Yes
Latex free (Yes/No)	NA	NA	Yes
Brand	Covidien	Covidien	Covidien
Manufacturer	Medtronic	Medtronic	Medtronic
Warranty (years)	3	3	3