

- · Lightweight, durable, rust-resistant frame
- Available with and without backrest
- Seat height adjusts in 1" increments for various user heights

• Wide seat with handles (7932 Only) provides additional comfort

Maximum Weight Capacity 600 lb - EVENLY DISTRIBUTED





Items 7932A & 7939A



Instructions for Use -Please Read All Instructions Before Use.

Setup

Installing the Back

(Only for 7939A)

- 1. Place bath seat on its side on a flat working surface.
- 2. Locate the channel guides (4 each) on underside of seat. Align backrest support tubes with channel guides.
- 3. Slide backrest seat support tubes through channel guides until seat support tube push buttons protrude fully through channel guide holes.

NOTE: Ensure seat support tube buttons protrude fully through channel guide holes.

Caution: Backrest is not designed to support the full weight of the user.

Adjusting Leg Height

Legs are already installed on the bath seat.

To adjust leg height

- 1. Press push button and slide leg extension up or down to desired adjustment hole.
- 2. Ensure push button fully protrudes through the leg adjustment holes of the leg extension.

NOTE: An audible "click" will be heard when push button securely protrudes through the adjustment hole.

3. Repeat STEPS 1 and 2 for the other three legs.

NOTE: Ensure all four legs are adjusted to the same height and that all height adjustment buttons protrude fully through the adjustment holes before use.



Maintenance

- Use a mild soap and water solution, rinse with clear water, then dry with a soft cloth. Ensure that product is well rinsed before using.
- Never use organic solvents, abrasive cleaners, or scouring pads on any part of the product.
- Check fit and tightness of all nuts, bolts and knobs weekly to ensure that product is safe
- To guarantee the best service from your product, always specify Lumex replacement parts, available from your GF Health Products, Inc. authorized distributor.

Safety Guidelines -**Please Read Before Use**

NOTE: Check ALL parts for shipping damage. If shipping damage is noted, DO NOT use. Contact carrier/dealer for further instruction.

WARNING

WARNING: DO NOT install or use this equipment without first reading and understanding these instructions. If you are unable to understand the Warnings, Cautions or Instructions, contact a healthcare professional, dealer or technical personnel before attempting to use this equipment - otherwise, injury or damage may occur.

WARNING: Item has a maximum weight capacity of 600 lb. EVENLY DISTRIBUTED

WARNING: GF Health Products, Inc. assumes no responsibility for any damage or injury caused by improper installation, assembly or use of this product.

WARNING: ALWAYS inspect the bath seat to ensure that it is properly positioned and stable before using. DO NOT use if bath seat is wobbly or unstable

WARNING: If components are damaged or missing, contact your GF Health Products, Inc. distributor immediately. DO NOT use substitute parts.

WARNING: Check leg tips for rips, tears, cracks or wear. If any of these conditions exist, replace leg tips IMMEDIATELY.

WARNING: DO NOT use this as a transfer bench or climbing device.

WARNING: All four leg tips MUST be in contact with shower/tub floor at all times.

WARNING: Ensure that the four chair legs are adjusted to the same height and that the push buttons fully protrude through their respective adjustment holes BEFORE use

WARNING: Only use GF Health Products, Inc. accessories with this product.

WARNING: This product should only be used with tub floors wider than 18 ¾ inches.

WARNING: Users with limited physical capabilities must be supervised or assisted when using the bath seat.

WARNING: Make certain ALL screws and nuts are tight at all times.

WARNING: Notice for California Customers
California Proposition 65 WARNING: This product contains a chemical known to the State of California to cause cancer and reproductive or developmental harm.

WARNING: DO NOT submerge product in water, otherwise damage may result.