

## INSTRUCTIONS FOR USE – PLEASE READ ALL INSTRUCTIONS BEFORE USE.

The Lumex 700175C-2 is designed as a dual purpose mobility aid. It is recommended for use by users with heights ranging from 5'4 to 6'2. Please consult a healthcare professional for the proper adjustment.

### SINGLE RELEASE FOLDING WALKER SIZE SELECTION CHART

Model Number	User Height	Adjustment Range/ Handgrip To Floor
700175C-2	5'4" - 6'2"	32" - 39"

Leg extensions are available for users over 6' 2".  
Please consult your Lumex dealer.

### HEIGHT ADJUSTMENT

*NOTE: For ease of shipping, the walker is shipped in its lowest height position. The walker should be adjusted to the desired height before operation.*

Refer To Illustration Below:

The legs of each walker are equipped with an adjustment button and an adjustable leg attachment with several adjustment positions.

1. Adjust the height of the walker by aligning all four leg attachments to the same position height.
2. To adjust the height of a leg attachment, depress the adjustment button and twist the leg attachment to the desired height.



### TO OPEN UNIT

1. Grasp one handgrip and pull away from center of walker.
2. Repeat with other handgrip.
3. Push handgrips away from each other until the release mechanisms click into place.
4. Walker is safely opened when push button in release mechanism is up.

### TO CLOSE UNIT

1. Push downward on release mechanism and hold. *NOTE: Sides of walker will collapse inward when release mechanism is depressed.*
2. Grab one handgrip and push toward center of walker.
3. Repeat for other handgrip.

### MAINTENANCE

- Wipe walker weekly with a damp cloth.
- Grips may be cleaned with mild soap or detergent. DO NOT use organic solvents, such as acetone, which may damage grip.
- Never use abrasive cleaners or scouring pads on any part of the walker.
- Check all hardware weekly to ensure that it is securely attached.
- Replace walker tips as needed.
- To guarantee the best service from your walker, always specify Lumex replacement parts, available from your Lumex dealer.

## SAFETY GUIDELINES – PLEASE READ BEFORE USE

1. This product should not be used without proper instruction from a healthcare professional.
2. This product has a maximum user weight limitation of 400 lb.
3. Before using, be sure that walker is locked in the open position.
4. Graham Field assumes no responsibility for any damage or injury caused by improper installation, assembly or use of this product.
5. If components are damaged or missing, contact your dealer immediately. DO NOT use substitute parts.
6. Inspect walker weekly for proper function and wear. If any component is not in proper working order or appears worn, contact your dealer for repair and/or appropriate replacement parts.
7. Before use, ensure that push button protrudes fully through height adjustment hole and the release mechanism is locked in the "open" position.
8. Always use caution when using this product on wet or slippery surfaces.