DisposableBelt

SystemRoMedic™

User manual – English

REF
6054

SWL: 100 kg/220 lbs

handicare

Manual no: 773 Ver. 11 150817

DisposableBelt

SystemRoMedic™

English

The DisposableBelt is intended for use as an accessory for extra support during various types of transfers and must not be used as an assistive device for lifting. The DisposableBelt is made of a very durable non-woven material. It must be discarded when soiled or when no longer required.

Check Safety

Visual inspection

Check the condition and function of the assistive device regularly. Always inspect the product after laundering. Check to ensure that seams and material are free from damage. Check to ensure that fabric is not worn or faded. If there are signs of wear, the product must be discarded.

Always read the user manual

Always read the user manuals for all assistive devices used during a transfer.
Keep the user manual where it is accessible to users of the product.
Always make sure that you have the right version of the user manual.
The most recent editions of user manuals are available for downloading from our website, www.handicare.com.

Placement



Apply the belt around the user, below the waist.

Select the appropriate size loop. Fasten the belt and tighten it.

Removing

Release the velcro and then remove the belt.

Using of DisposableBelt

1

2

3

4

Caregivers can work in pairs or alone, depending on the user's needs, since DisposableBelt have several vertical and horizontal handles. Cross-grips at the back provide support and enable a good working stance (1 and 2).

It is important to lean the user forward to get a natural pattern of movement, working in towards the body with DisposableBelt.

When two caregivers are assisting, it is appropriate if one caregiver stands on each side, grasping the cross-grips at the back and holding the user by the hand at the front (3).

DisposableBelt provide good support during transfers between wheelchair and bed (4).

Material

Polypropylene

Care of the product

Read the product label.

Do not launder. Discard the product if it is soiled or if there is a need to use another product.

CE Medical Device Class I

Medical Device Class I. The product complies with the requirements of the Medical Devices Directive 93/42/EEC.

Product information



Size	cm	inch
Medium	84-94	33-37
Large	92-102	36-40